# A41I-2745: Advantages of Hourly Resolved Formaldehyde (HCHO) Measurements in Identifying Impacts from Wildfire Emissions, a Case Study.

## Karsten Baumann<sup>1,2</sup>, J.C. Guerrero<sup>1</sup>, Jonathan Bent<sup>1</sup>, Gregor Lucic<sup>1</sup>, Kai Skog<sup>1</sup>, Chris Rella<sup>1</sup>, and Joel Avrunin<sup>1</sup>

<sup>1</sup>Picarro, Inc., Santa Clara, CA 95054; <sup>2</sup>Dept. of Environmental Sciences and Engineering, Gillings School of Global Public Health, UNC-Chapel Hill, Chapel Hill, NC 27599


### 0. OBJECTIVE

Data obtained from the Picarro G2307. a Cavity Ring Down Spectrometer (CRDS) for formaldehyde (HCHO) has previously demonstrated excellent overall correlation with data obtained via EPA Method TO-11A. Extensive intercomparison with trusted standards, and key opinion leader laboratory instruments, and a traceable calibration process has led to a revision of the instrument scaling that matches TO-11A methods very closely and consistently. To further improve performance, we introduced auto-zero functionality, allowing to track the instrument zero baseline in a highly resolved manner. We believe, by correcting for its drift over time, a lower detection limit and improved accuracy in different ambient outdoor and indoor air matrices are achievable.


- Manufacturing

- CH₄ + OH methane
  - methanol
  - - isoprene

### 2. MEASUREMENT LOCATION "RTP"







### **3. INSTRUMENTATION**


Picarro developed the G2307 HCHO instrument between 2017 and 2018 in response to a need for stable real-time instrument was equipped with an adjacent  $CH_4$  line to enable both real-time  $CH_4$  measurements and an ability to determine laser and system health without needing to directly check/calibrate with HCHO.

G2307 Instrument (near-IR CRDS)		PICARRO
G2307 Formaldehyde	Typical Performance**	Specifications***
Lower Detection Limit (3 <i>o</i> , 300 sec)	0.18 ppb	0.3 ppb
<b>Zero Drift</b> (72 hrs)* (peak-to-peak, 50-minute average)	0.33 ppb	1.5 ppb
Precision (1 $\sigma$ , 2 sec) Precision (1 $\sigma$ , 10 sec) Precision (1 $\sigma$ , 300 sec)	0.7 ppb + 0.1% of reading 0.31 ppb + 0.05% of reading 0.06 ppb + 0.02% of reading	1.2 ppb + 0.1% of reading 0.6 ppb + 0.05% of reading 0.1 ppb + 0.02% of reading
Measurement Interval	<2 sec	<2 sec
Response Time (0–20 ppb) (Rise/Fall Time 10–90% / 90–10%)	<5 sec	<5 sec
Measurement Range	0–30 ppm	0–30 ppm
* Picarro analyzers do not require a zero reference gas or zero catridge to operate or meet specifications.	** Typical performance is defined as the median of testing results from ten sequentially built G2307 analyzers in 2019. Results available upon request.	*** Specifications and an instrument-specific testing report (Certificate of Compliance) provided with every analyzer purchase.

### 4. SETUP

Goal: Outdoor vs Indoor with frequent zeros and standard additions at either Outdoor Inlet to check

PICAR RO





**GILLINGS SCHOOL OF GLOBAL PUBLIC HEALTH**